

**Requirements for Calibration Service Providers**

Calibrations shall be performed by service providers meeting one of the following requirements: ISO 17025: 2017, ANSI/NCSL Z540.

As a part of the calibration service being contracted labs are required to provide certificates (and labels if required) for the timeframes indicated below, corresponding to the gage type being requested.

<b>Type Gage</b>	<b>Calibration Certificate Requirement</b>
Thread Plug Gage	3 years
Thread Set Plug Gage	3 years
Ring Thread Gage	2 years
Pipe Thread Ring Gage	3 years
Pipe Thread Plug Gage	3 years
Calipers	1 year
Micrometers	2 years
Counting Scales	1 year
Gage Block Set	5 years
Height Gage	3 years
Height Gage Dial Indicator	3 years
Surface Plate	5 years
Optical Comparator	3 years
Torque Wrench	1 year
Pin Gage	3 years
Temperature Gage	2 years
Light Meter	2 years